



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/606,303	06/25/2003	David Wayne Daniels	DEP651	8037
27777	7590	10/05/2005	EXAMINER	
PHILIP S. JOHNSON JOHNSON & JOHNSON ONE JOHNSON & JOHNSON PLAZA NEW BRUNSWICK, NJ 08933-7003			ARAJ, MICHAEL J	
			ART UNIT	PAPER NUMBER
			3732	

DATE MAILED: 10/05/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No.	Applicant(s)	
	10/606,303	DANIELS ET AL.	
	Examiner	Art Unit	
	Michael J. Araj	3732	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 15 September 2005.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,2,4,5,7-17,19 and 20 is/are rejected.
- 7) ☒ Claim(s) 3,6 and 18 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 25 July 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|--|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input checked="" type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date <u>7/25/2003</u> | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Election/Restrictions

Applicant's election without traverse of Species 1 of the modular reamer and Species I of the hip stem trial in the reply filed on September 15, 2005 is acknowledged.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1,2,4,5, 7-17 and 19 rejected under 35 U.S.C. 102(e) as being anticipated by McGovern (U.S. Publication No. 2004/0147933).

McGovern discloses a kit or system comprising hip femoral component trials (400) with different dimensions, a reamer (102) with a tapered external periphery having a first portion (100) adapted for placement at least partially in the cavity of the long bone and a second portion (200) operably connected and a second portion removably connectable to a driver (300) that rotates said reamer (102), wherein the driver(300) defines a longitudinal axis and the reamer defines another longitudinal axis that are coincident and separable along the same axis (110). Also disclosed is a locking feature associated with at least one of the trials and reamer for locking at least one of the drivers and trials to said reamer (page 3, paragraph 47, lines 7-11). McGovern also

discloses a method for providing joint arthroplasty comprising resecting a long bone e.g. femur, opening a medullar canal of the long bone, providing a reamer including a surface, attaching a driver to the reamer, positioning the reamer in the canal, reaming a cavity in the canal with the reamer, detaching the driver from the reamer, providing a trial, attaching the trial to the reamer, performing a trial reduction, removing the reamer and the trial, providing a joint prosthesis and implanting the joint prosthesis in the cavity.

With regard the statement of intended use and other functional statements, they do not impose any structural limitations on the claims distinguishable over McGovern which is capable of being used as claimed if one so desires to do so. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963).

Furthermore, the law of anticipation does not require that the reference "teach" what the subject patent teaches, but rather it is only necessary that the claims under attack "read on" something in the reference. *Kalman v. Kimberly Clark Corp.*, 218 USPQ 781 (CCPA 1983). Furthermore, the manner in which a device is intended to be employed does not differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. *Ex parte Masham*, 2 USPQ2d 1647 (1987).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 20 is rejected under 35 U.S.C. 103(a) as being unpatentable over Littenberg (U.S. Publication No. 2004/0147933) in view of Scifer et al. (U.S. Publication No. 2003/0171816).

Littenberg discloses the claimed invention except for a driver having indicia thereon corresponding to at least one of the bony landmarks on a patient and a portion of a trial or implant. Scifert discloses marks (60) on the trial stem for height positioning with respect to the surface of the long bone. It would have been obvious to one skilled in the art at the time the invention was made to construct the driver of Littenberg with markings in view of Scifert, for height positioning of the implant. Though, Scifert discloses the marking on the implant rather than the driver, both are functionally equivalent.

Allowable Subject Matter

Claims 3, 6 and 18 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See PTO-892 for art cited of interest.


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael J. Araj whose telephone number is 571-272-5963. The examiner can normally be reached on M-F 8am-5pm.

Art Unit: 3732

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).


MJA


EDUARDO C. ROBERT
PRIMARY EXAMINER